

Pacemaker lead disposition in an elderly patient - Reel syndrome

Farhat Fouladvand, Oktaj Maksudov

UMBAL Heart and Brain hospital Burgas, Bulgaria, Bulgarian Cardiac Institute

We present the case of an adult patient with Reel syndrome who had a VVI-mode permanent pacemaker implanted 6 months earlier. X-ray revealed the retraction and winding of the right ventricle chamber electrode around the pulse generator. Subsequently, a new electrode was reimplanted in the right ventricle apex.

Key words: Reel syndrome, pacemaker